

Using 10 cm Novex® and NuPAGE® Mini Gels with the Bolt® Mini Gel Tank

Cat. no. B4477599


Publication No. MAN0008925

Rev. B.0


Instructions for using 10 cm Novex® and NuPAGE® Mini Gels with the Bolt® Mini Gel Tank fitted with **black cam handles** for 10.5 cm mini gels are described below.

Samples, running buffers, mini gels, and the Bolt® Mini Gel Tank should be prepared normally as directed in the standard instructions. For detailed protocols, or instructions for changing black cam handles for 10.5 cm mini gels to gray cam handles for 10 cm mini gels refer to the applicable manuals available from www.lifetechnologies.com/manuals.


Procedure



1. Place the cassette into the chamber with the wells facing towards you (short plate facing front).



2. Make sure the gel cassette is centered/aligned, and flush with the bottom of the chamber so a proper seal is formed with the gasket.
3. Hold the gel cassette in its centered position, and secure the gel cassette by pulling the cam handle forward.



4. Fill the Anode and Cathode Chambers with 1X Running Buffer. For reduced samples, use 200 mL 1X Running Buffer with 0.5 mL NuPAGE® or Bolt® Antioxidant in the Cathode Chamber (and Overflow Compartment). Make sure the loading wells are completely filled with buffer (the Overflow Compartment should be nearly full).
5. Load your samples and markers.
6. Perform electrophoresis using the following conditions:

Gel Type	Voltage (constant)	Run Time
Novex® Tris-Glycine Mini Gel	125 V	100 minutes
NuPAGE® Bis-Tris Mini Gel (with MES Buffer)	200 V	30 minutes
NuPAGE® Bis-Tris Mini Gel (with MOPS Buffer)	200 V	42 minutes
Bolt® Bis-Tris Plus Mini Gel (with MES Buffer)	200 V	20 minutes
Bolt® Bis-Tris Plus Mini Gel (with MOPS Buffer)	200 V	35 minutes
NuPAGE® Tris-Acetate Mini Gel	150 V	50 minutes
Novex® Tricine Mini Gel	125 V	65 minutes

Run time may vary depending on gel percentage and the power supply used for electrophoresis.

Limited Product Warranty

Life Technologies Corporation and/or its affiliate(s) warrant their products as set forth in the Life Technologies' General Terms and Conditions of Sale found on Life Technologies' website at www.lifetechnologies.com/termsandconditions.

If you have any questions, please contact Life Technologies at www.lifetechnologies.com/support.

Important Licensing Information: This product may be covered by one or more Limited Use Label Licenses. By use of this product, you accept the terms and conditions of all applicable Limited Use Label Licenses.

DISCLAIMER: LIFE TECHNOLOGIES CORPORATION AND/OR ITS AFFILIATE(S) DISCLAIM ALL WARRANTIES WITH RESPECT TO THIS DOCUMENT, EXPRESSED OR IMPLIED, INCLUDING BUT NOT LIMITED TO THOSE OF MERCHANTABILITY, FITNESS FOR A PARTICULAR PURPOSE, OR NON-INFRINGEMENT. TO THE EXTENT ALLOWED BY LAW, IN NO EVENT SHALL LIFE TECHNOLOGIES AND/OR ITS AFFILIATE(S) BE LIABLE, WHETHER IN CONTRACT, TORT, WARRANTY, OR UNDER ANY STATUTE OR ON ANY OTHER BASIS FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH OR ARISING FROM THIS DOCUMENT, INCLUDING BUT NOT LIMITED TO THE USE THEREOF.

© 2014 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified.

For support visit www.lifetechnologies.com/support or email techsupport@lifetech.com

www.lifetechnologies.com

24 September 2014

